


DUV Raman/PL 200



Fully Integrated,
Lab Model
Deep UV Resonance Raman &
Photoluminescence
Spectrometer,
with microscopic imaging

Features

Why Deep UV Spectroscopy?

Deep UV excited Raman & photoluminescence spectroscopy is an emerging analytical instrument technology with vast potential for a wide range of commercial, industrial, and research applications. A major limitation of Raman spectroscopy conducted in the near UV, visible, or near IR is obscuration or interference of the Raman signals due to background fluorescence from the analyte or its background or surroundings within the laser beam interrogation spot. This interference limits the types of materials and compositions or backgrounds for which Raman spectroscopy is useful.

Native fluorescence is a phenomenon which does not occur below about 270 nm for the vast majority of materials, independent of excitation wavelength. Raman, on the other hand, is dependent on excitation wavelength and when excitation occurs below 250 nm, there exists a spectral region within which to observe over 3000 cm^{-1} of Raman shifted emissions without obscuration or interference from fluorescence.

This is the driving motivation for spectroscopy conducted in the deep UV with excitation below 250 nm. Also, operation in the deep UV enables simultaneous detection of Raman and fluorescence spectra, enabling the much higher sensitivity of fluorescence and the higher specificity of Raman spectra. There are many other advantages of operation in the deep UV including enhanced Raman signal strength due to resonance effects, simplification of Raman spectra for resonant materials making spectra easier to interpret, small depth of penetration into many materials, which limits interference with background or substrate materials, and other benefits.

Deep UV resonance Raman & photoluminescence spectroscopy has been hampered by the lack of a suitable laser source in the deep UV. Photon Systems has developed a new enabling laser technology to address this problem. These lasers have exceptionally narrow and stable emission linewidths and are hundreds of times smaller, lighter, and lower power consumption than other deep UV lasers. Photon Systems lasers have been vetted in a wide range of harsh commercial, industrial, and research and have been selected by NASA for the Mars 2020 lander mission for a rover arm mounted deep UV Raman and fluorescence instrument.

Excitation Wavelength: 248.6nm.

Polarization: Random (Raman signatures insensitive to polarization of target materials)

Spectrograph: 200 cm Czerny Turner with dual computer controlled 4200 & 600 g/mm holographic gratings

Dispersion: 1.9 $\text{cm}^{-1}/\text{pixel}$ (w 4200 g/mm grating)

Resolution: $<8 \text{ cm}^{-1}$, 75 μm slit

Entrance Slits: fixed, selectable

Spectral Spread: 400-4400 cm^{-1} (4200 g/mm grating)
250nm to 650 nm (600 g/mm grating) (0.6 nm res)

Detector: 3 stage TE Cooled, back illuminated UV CCD Array

Obj. Lens: 3X, 5X, 15X, 40X DUV achromatic objectives (10 mm entrance aperture diameter)

Context Imaging Camera: 2.4 M pixel

FOV: 1.3mm, 267 μ , 100 μ

Motorized Position/Mapping Stage: 5 x 5 cm mapping area, optional as shown.

Overall Size: 7.0" W x 7.3" H x 15" D

Weight: $<20 \text{ lb}$

Power Consumption:

Standby 8 W

Max power 60 W

Input: 85VAC to 270VAC
or 24 VDC

Safety: Class I, DHHS/CDRH

Command & Control: External laptop computer for command, control, data processing, chemometrics, data storage, etc.

Patents: U.S. Patents 6,278,869, 7,800,753, 8,395,770, plus pending.

